OCTOPI™
FIXATION EVOLVED

FIELD ORTHOPAEDICS

TECHNICAL SPECIFICATIONS

OCTOBridge™	1.4 OCTOBridge™	1.7 OCTOBridge™	2.4 OCTOBridge™
K-wire size	1.4mm	1.0mm	1.2mm
Drill size	1.4mm	1.75mm	2.4mm
Drill Depth	12mm	12mm	17mm
Suture USP	#2-0	#0	#2
No. of Needles	2		
Needle type	15mm, 3/8 circle	22mm, 3/8 circle	22mm, 3/8 circle

OCTOExtend™	1.7 OCTOExtend™	2.4 OCTOExtend™
K-wire size	1.0mm	1.2mm
Drill size	1.75mm	2.4mm
Drill Depth	12mm	17mm
Suture USP	#2-0	#2-0
No. of Needles	N/A	
Needle type	N/A	

OCTOAnchor™	1.4 OCTOAnchor™	1.7 OCTOAnchor™	2.4 OCTOAnchor™
K-wire size	1.4mm	1.0mm	1.2mm
Drill size	1.4mm	1.75mm	2.4mm
Drill Depth	12mm	12mm	17mm
Suture USP	#2-0	#0	#2
No. of Needles	4		
Needle type	15mm, 3/8 circle	22mm, 3/8 circle	22mm, 3/8 circle

This material is intended for health care professionals. Distribution to any other recipient is prohibited. For product information, including indications, contraindications, warnings, precautions, potential adverse effects, consult the Instructions For Use at www.fieldorthopaedics.com/brochures. Each surgeon should exercise his or her own independent judgment in the diagnosis and treatment of an individual patient, and this information does not purport to replace the comprehensive training surgeons have received. As with all surgical procedures, the technique used in each case will depend on the surgeon's medical judgment as the best treatment for each patient. Not all patients are candidates for this product and/or procedure. Products listed may not be available in all markets. Please contact Field Orthopaedics if you have any questions about the availability of products in your area.

MANUFACTURER
Field Orthopaedics Pty. Ltd.
30 Florence St.
Teneriffe QLD 4005 Australia

sales@fieldorthopaedics.com
www.fieldorthopaedics.com
TS00003 v3 Nov 2026